



International Journal of Medicine and Pharmaceutical Research

Journal Home Page: www.pharmaresearchlibrary.com/ijmpr



Research Article

Open Access

A New RP-HPLC Method Development and Validation for the Amlodipine and Perindopril in Bulk and Tablet Dosage Form

Y. Krishna Reddy*, A. Harikrishna, B. Tharun, B. Kishore, B. Saidhulu, B. Rajitha

Bomma Institute of Pharmacy, Khammam, Telangana, India

ABSTRACT

The chromatographic conditions were successfully developed for the separation of Amlodipine and Perindopril by using Kromosil C 18 (250×4.6mm, 5μ) Column, The flow rate was 1ml/min, mobile phase ratio ortho phosphoric acid buffer: Methanol 65:35 were set (Buffer P^H 2.45 adjusted with Triethylamine). The instrument used was Shimadzu UFLC-20 AD Chromatographic system (Japan), SPD-M20A diode array, LC 20 software. The optimum wavelength for detection was 254 nm at which better detector response for both the drugs was obtained. The retention times for Amlodipine and Perindopril was found to be 2.589 ± 0.004 min and 3.711 ± 0.005 min, respectively. To ascertain its effectiveness, system suitability tests were carried out on freshly prepared stock solutions. The calibration was linear in concentration range of 20 to 60 μg/ml and 10 to 30 μg/ml, with regression 0.9979 and 0.9999, Amlodipine and Perindopril respectively. The low values of % R.S.D indicate the method is precise and accurate. The mean recoveries were found above 99.3 % for both the drugs. Robustness of the proposed method was determined by varying various parameters, the %RSD reported was found to be less than 2 %. The proposed method was validated in accordance with ICH parameters and the applied for analysis of the same in marketed formulations.

Keywords: Amlodipine, Perindopril, HPLC

ARTICLE INFO

CONTENTS

1. Introduction	191
2. Materials and Methods	192
3. Results and discussion	193
4. Conclusion	196
5. References	196

Article History: Received 25 May 2016, Accepted 29 June 2016, Available Online 10 August 2016

*Corresponding Author

Y. Krishna Reddy
Bomma Institute of Pharmacy,
Khammam, Telangana, India
Manuscript ID: IJMPR3046



PAPER-QR CODE

Citation: Y. Krishna Reddy, et al. A New RP-HPLC Method Development and Validation for the Amlodipine and Perindopril in Bulk and Tablet Dosage Form. *Int. J. Med. Pharm. Res.*, 2016, 4(4): 191-197.

Copyright©2016 Y. Krishna Reddy, et al. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution and reproduction in any medium, provided the original work is properly cited.

1. Introduction

Methods are developed for new products when no official methods are available. Alternate methods for existing (non-

pharmacopoeial) products are developed to reduce the cost and time for better precision and ruggedness [1]. Trial runs

are conducted, method is optimized and validated. When alternate method proposed is intended to replace the existing procedure comparative laboratory data including merit/demerits are made available [2].

Description of the Various Analytical Methods

Titrimetric and gravimetric method of analysis is suitable when the sample is present in pure form or when no interference is observed in the mixture with other materials [3]. Ultraviolet and visible spectrometric method is suitable when no Interference is observed in the mixture[4]. HPLC and GC methods are more advantageous than the above due to their capability in separating organic mixtures and quantitative estimations. AAS is used mainly for quantitative estimation in ppm and ppb levels of elements [5]. Infra-red spectroscopy though mainly used for qualitative analysis can be used for quantitative estimation also. Out of all the above methods, thin layer chromatography plays a very important role in analysis due to its adaptability, flexibility, and cost and time. It can be used both for qualitative and quantitative determination. After separation spots can be scanned with the help of a scanner and quantitative measurement can be made [6].

Chromatography:

Chromatography is a technique used in analytical chemistry to separate and identify components of mixtures. The name comes from the Greek term for "color writing" because this method was originally used to separate colored samples. The advent of high-performance liquid chromatography (HPLC).in this system pressure is applied to the column, forcing the mobile phase through at much higher rate[7]. The pressure is applied using a pumping system. The action of the pump is critical, since it must not pulsate and mix up the sample being separated in the solvent, causing it to lose resolution [8]. Development of pumps has proceeded quite quickly over the last several years, and now it is possible to achieve good resolution under the conditions required for HPLC [9].




Figure 1: Amlodipine




Figure 2: Perindopril

2. Materials and Methods

Apparatus: The instrument used for the study was Shimadzu UFLC-20 AD Chromatographic system (japan), SPD-M20A diode array, LC 20 software

Reagents and Materials The solvents used were Potassium dihydrogen orthophosphate, Methanol, Triethylamine, Orthophosphoric acid and Water[10].

Selection of detection wavelength:

The sensitivity of method that uses UV- Vis detector depends upon the proper selection of wavelength. An ideal wavelength is that gives maximum absorbance and good response for both the drugs to be detected. Standard solutions of Amlodipine and Perindopril were scanned in the UV range (200-400nm) and the spectrums obtained were overlaid and the overlain spectrum was recorded. From the overlain spectrum, 254 nm was selected as the detection wavelength for the present study [11].

Selection of mobile phase

Initially the mobile phase tried was methanol and water, methanol and Methanol, buffer and water in various proportions. Finally, the mobile phase was optimized to Buffer: Methanol in proportion 65:35 v/v respectively [12].

Optimization chromatographic conditions

Optimization Chromatographic trials for Simultaneous Estimation of Amlodipine and Perindopril by RP-HPLC

Parameters Description

Column: kromosil C18 Column (100mm x 4.6mm) 5µg.
 Mobile phase ratio: Methanol: Phosphate buffer PH 2.5 (35:65 v/v)
 Detection wavelength: 254 nm
 Flow rate: 1ml/min Injection volume :
 20µl Column temperature: Ambient Auto sampler
 Temperature: Ambient




Figure 3: Optimization Chromatogram

Observation: The separation of two analytical peaks was good. The plate count also above 2000, tailing factor below 2, and the resolution is above 2. The condition is taken as optimized method.

Procedure

Preparation of Buffer:

About 7.0g of potassium dihydrogen orthophosphate was dissolved in 1000ml of HPLC grade water and pH 2.5 was adjusted with orthophosphoric acid. It was filtered through 0.45µm nylon membrane filter and degassed with sonicator. It was used as a diluent for the preparation of sample and standard solution [13].

Preparation of mobile phase:

Mobile phase consist of buffer: Methanol of P^H 2.5 (35:65) was taken sonicated and degassed for 10min and filtered through 0.45 µm nylon membrane filter

Standard Preparation:

Weigh accurately 10 mg Amlodipine Working Reference Standard and 15mg of Perindopril Working Reference Standard is taken in to 100ml volumetric flask and then it was dissolved and diluted to volume with mobile phase up

to the mark. After that 50ml of the above solution was taken into 100ml standard flask and made up with mobile phase. (Stock solution) Further pipette 0.5ml of the above stock solution in to a 10ml volumetric flask and dilute up to the mark with diluents [14].

Assay

Preparation of samples for Assay

Standard preparation:

Weigh accurately 10mg Amlodipine Working Reference Standard and 15mg of Perindopril Working Reference Standard is taken in to 100ml volumetric flask and then it was dissolved and diluted to volume with mobile phase up to the mark. After that 50ml of the above solution was taken into 100ml standard flask and made up with mobile phase. (Stock solution). Further pipette 0.5ml of the above stock solution in to a 10ml volumetric flask and dilute up to the mark with diluents [15].

Sample preparation:

10 tablets were weighed and calculate the average weight of each tablet then the weight equivalent to 10 tablets was transferred into a 100ml standard flask. A volume of 70ml of mobile phase was added and sonicate for 30min. Then the solution was cooled and diluted to volume with mobile phase and filtered through 0.45µm membrane filter [16]. (Stock solution). Further pipette 0.25ml of Amlodipine and Perindopril of the above stock solution in to a 10ml volumetric flask and dilute up to the mark with diluent.

Assay procedure


20µL of the standard and sample solutions of Amlodipine and Perindopril were injected into the HPLC system and the chromatograms were recorded. Amount of drug present in the capsules were calculated using the peak areas [17].

3. Results and Discussion


Method Validation Parameters

1. Specificity

ICH defines specificity as “the ability to assess unequivocally the analyte in the presence of components which may be expected to be present. Typically this might include impurities, degradants, matrix, etc.



Chromatogram of Blank




Chromatogram of Sample


2. Linearity

The linearity of an analytical method is its ability to elicit test results that are directly/ by a well-defined mathematical transformation, proportional to the concentration of analyte in samples within a given range. Serial dilutions of Amlodipine and Perindopril (20-60µg/ml and 10-30 µg/ml) were injected into the column and detected at a wavelength set at 254 nm. The calibration curve was obtained by plotting the concentration vs. peak area [18].

Acceptance criteria: Correlation coefficient should be not less than 0.999.



Calibration graph of Amlodipine



Calibration graph of Perindopril

3. Range

The linearity study was performed for the concentration of 20ppm to 60ppm and 10ppm to 30ppm level. Each level was injected into chromatographic system. The area of each level was used for calculation of correlation coefficient.

4. Accuracy

Accuracy of the method was determined by recovery experiments. There are mainly 2types of recovery studies are there.

- a) Standard addition method: To the formulation, the reference standard of the respective drug of known concentration was added, analyzed by HPLC and compared with the standard drug concentration.
- Percentage method: For these assay method samples are prepared in three concentrations of 50%, 100%, and 150% respectively[19].

Acceptance criteria: The mean % recovery of the Amlodipine and Perindopril at each level should be not less than 95.0% and not more than 105.0%.

5. Precision

The precision of the method was demonstrated by intra-day and inter-day precision studies. Intra-day studies were performed by injecting three (3) repeated injections within a day. Peak area and %RSD were calculated and reported. The chromatograms of intra-day precision studies were shown. Inter-day precision studies, was done by injecting three (3) repeated injections for three consecutive days. Peak area and %RSD were calculated and reported [20].

Method Precision:

Method precision also called as repeatability/Intra-day precision indicates whether a method gives consistent results for a single batch. Method precision was demonstrated by preparing six test solutions at 100% concentration as per the test procedure & recording the chromatograms of six test solutions [21]. The % RSD of peak areas of six samples was calculated. The method

precision was performed on Amlodipine and Perindopril formulation. The % RSD of the assay value for six determinations should not be more than 2.0%.

Intermediate Precision:

Intermediate precision of the analytical method was determined by performing method precision on another day by different analysts under same experimental condition. Assay of all six replicate sample preparations was determined and mean %assay value, standard deviation & %RSD was calculated [22].

Detection limit

The LOD was performed for Amlodipine and Perindopril was found to be 0.001 and 0.005 respectively.

Quantitation limit

The LOQ was performed for Amlodipine and Perindopril was found to be 0.004 and 0.015 respectively.

Table 1: Calibration data of Amlodipine and Perindopril

Sample ID	Amlodipine		Perindopril	
	Concentration (mcg/ml)	Area	Concentration (mcg/ml)	Area
20% of operating concentration	20	1224140	10	740046
40% of operating concentration	30	1595681	15	990204
60% of operating concentration	40*	1992966	20*	1183023
80% of operating concentration	50	2356546	25	1439886
100% of operating concentration	60	2797214	30	1682302
Correlation Coefficient			0.999	

Table 2: Accuracy results for Amlodipine

Sample Id	Conc found (µg/ml)	Concn Obtained (µg/ml)	%Recovery	Mean recovery	Statistical Analysis
50%	5	5.01	100.2		%RSD= 0.505
50%	5	4.96	99.2	99.73	
50%	5	4.99	99.8		
100%	10	9.95	99.5		%RSD=0.66
100%	10	9.87	98.7	98.8	
100%	10	9.82	98.2		
150%	15	14.64	97.6		%RSD=1.45
150%	15	14.76	98.4	98.8	
150%	15	15.06	100.4		

Table 3: Accuracy results for Perindopril

Conc (µg/ml)	Concn Obtained(µg/ml)	%Recovery of drug	Mean accuracy	%RSD
5	4.92	98.0	99.2	1.2
5	4.96	99.2		
5	5.02	100.4		
10	9.95	99.5	99.5	0.2
10	9.94	99.4		
10	9.98	99.8		
15	14.78	98.6	99.0	0.530
15	14.94	99.6		
15	14.83	98.8		

Table 4: Robustness data for Amlodipine

Std. Replicate	Variation in flow rate		Variation in Mobile phase composition	
	Flow Rate 0.8ml/min	Flow Rate 1.2ml/min	Buffer: Methanol (40:60)	Buffer: Methanol (30:70)
1	2492492	1676589	1951632	1979168
2	2495874	1675428	1954783	1967452
Mean	2494183	1676009	1953208.0	1973310
SD	2391.4	820.9	2228.0	8284.46
%RSD	0.09	0.04	0.11	0.4
Retention time	3.150	2.168	2.618	2.572
Tailing factor	1.4	1.3	1.3	1.3
Theoretical plates	5752	4207	4577	4476

Table 5: Robustness data for Perindopril

Parameter	Variation in flow rate		Variation in Mobile phase composition	
	Flow Rate 0.8ml/min	Flow Rate 1.2ml/min	Buffer: Methanol (40:60)	Buffer: Methanol (30:70)
1	1500192	100524	1196996	1153397
2	1500426	100468	1198547	1154782
Mean	1500309	100496	1197772	1154090
SD	165.5	39.59	1096.2	979.34
%RSD	0.01	0.03	0.09	0.08
Retention time	4.674	3.121	4.394	3.331
Tailing factor	1.2	1.2	1.2	1.2
Theoretical plates	7187	5412	6498	6471

Table 6: Method Precision data for Amlodipine & Perindopril

S.No	Concentration (µg/ml)	Amlodipine		Perindopril	
		Retention time (Rt)	Peak Area	Retention time(Rt)	Peak Area
1	40 & 20	2.586	2010800	3.713	1184689
2	40 & 20	2.588	2002956	3.714	1188199
3	40 & 20	2.590	2012800	3.734	1195842
4	40 & 20	2.590	2005243	3.737	1184210
5	40 & 20	2.591	2011092	3.741	1198327
6	40 & 20	2.589	2011098	3.740	1198320
Avg			2008998		1191598
SD			3920.9		6668.5
%RSD			0.19		0.55

Table 7: Intermediate Precision data for Amlodipine and Perindopril

S.No.	Concentration (µg/ml)	Intermediate Precision			
		Day 1 Amlodipine		Day 1 Perindopril	
		Retention time	Peak Area	Retention time	Peak Area
1	40&20	2.591	2005053	3.741	1183951
2	40&20	2.590	2007362	3.734	1184689
3	40&20	2.590	2007473	3.737	1186232
4	40&20	2.586	2009153	3.714	1186406
5	40&20	2.583	2012800	3.713	1188564
6	40&20	2.590	2012785	3.737	1187621
Avg			2009104		1186244
SD			3140.6		1730.9
%RSD			0.15		0.14

Table 8: LOD and LOQ Data of Amlodipine and Perindopril

Amlodipine			Perindopril		
Conc.(x) ($\mu\text{g/ml}$)	Peak Areas (y)	Statistical Analysis	Conc.(x) ($\mu\text{g/ml}$)	Peak Areas (y)	Statistical Analysis
40	2004682	S = 39092 c = 618048 LOD: 0.001 $\mu\text{g/ml}$ LOQ: 0.004 $\mu\text{g/ml}$	20	1184227	S = 39092 c = 369381 LOD: 0.005 $\mu\text{g/ml}$ LOQ: 0.015 $\mu\text{g/ml}$

4. Conclusion

The proposed HPLC method was found to be simple, specific, precise, accurate, rapid and economical for simultaneous estimation of Amlodipine and Perindopril in tablet dosage form. The developed method was validated in terms of accuracy, precision, linearity, robustness and ruggedness, and results will be validated statistically according to ICH guidelines. The Sample recoveries in all formulations were in good agreement with their respective label claims. From literature review and solubility analysis initial chromatographic conditions Mobile phase ortho phosphoric acid buffer: Methanol 65:35 were set (Buffer pH 2.45 adjusted with Triethylamine), Kromosil C 18 (250 \times 4.6mm, 5 μ) Column, Flow rate 1.0 ml/min and temperature was ambient, eluent was scanned with PDA detector in system and it showed maximum absorbance at 254 nm. As the methanol content was increased Amlodipine and Perindopril got eluted with good peak symmetric properties. The retention times for Amlodipine and Perindopril was found to be 2.589 min and 3.711 min respectively. System suitability parameters were studied by injecting the standard five times and results were well under the acceptance criteria. Linearity study was carried out between 50% to 150 % levels, R^2 value was found to be as 0.999. by using above method assay of marketed formulation was carried out, 100.7% was present. Full length method was not performed; if it is done this method can be used for routine analysis of Amlodipine and Perindopril.

5. References

- [1] Douglas A.Skoog, F. James Holler & Stanley R. Crouch. Instrumental analysis, India edition, 2007, pg: 13-14.
- [2] Gurdeep R. Chatwal & Sham K. Anand. Instrumental Methods Of Chemical Analysis (Analytical Chemistry) , pg: 2.566-2.567.
- [3] Ahuja S & Dong MW. Handbook of Pharmaceutical Analysis by HPLC. 1st edition, Academic Press Publisher.UK 2005.
- [4] Willard HH, Merrit LL, Dean JA, Settle FA. Instrumental methods of analysis, CBS Publishers and Distributors, New Delhi, 6th edition, 1986, 1-15.
- [5] Douglas A. Skoog, F. James Holler, Timothy A. Nieman. Principles of instrumental analysis, Saunders Golden Sun burst Series, Philadelphia, 2nd edition, 1980, 725-760.
- [6] David G.Watson. Pharmaceutical Analysis, A text book for Pharmacy students and Pharmaceutical Chemists, Harcourt Publishers Limited, 2nd Edition, 1999, 221-232, 267-311.
- [7] Snyder LR, Kirkland JJ, Joseph LG. Practical HPLC Method Development, Wiley Inter Science, New York, 2nd Edition, 1997, 1-56, 234-289, 685-712.
- [8] Satinder Ahuja & Neil Jespersen. Modern Instrumental Analysis 47 (Comprehensive Analytical Chemistry) volume-47, pg-7-8.
- [9] Beckett A.H, J.B. Stenlake. Practical Pharmaceutical Chemistry, 4th edition. C.B.S. Publications, Pg. No.53-62.
- [10] Remington's The Science and Practise of Pharmacy, 20th Edition, 2000.
- [11] Connors KA. A Textbook of Pharmaceutical Analysis, Wiley intersciences Inc, New Delhi, 3rd Edition, 1994, 373-421.
- [12] Rashmin. B. Patel, Mrunali R. Patel, An Introduction to Analytical method development for pharmaceutical formulations, Pharmainfo.net 2008; 17:19
- [13] Sharma B.k Instrumental methods of chemical analysis. 19 ed: Goel Publishing House, 2003.
- [14] Galen Wood Ewing, Instrumental methods of chemical analysis, 340-345.
- [15] United States of Pharmacopeia, USP30-NF25, The official compendia of standards, official May 1, 2007.
- [16] ICH topic Q2B, validation of analytical procedure & methodology, The European agency for evaluation of medicinal products, human medicines evaluation unit 1996.
- [17] ICH: Q2A, Text on validation of analytical procedure (October 1994).
- [18] Dipti B. Patel, N. J. Patel, S. K. Patel, A. M. Prajapati, and S. A. Patel, Rp hplc method for the estimation of Dutasteride in tablet dosage form. Indian Journal of Pharmaceutical Sciences, 2010 Jan-Feb; 72(1): 113–116.
- [19] Kamepalli Sujana, D. Gowri Sankar, Konda Abbulu and O.Bala Souri, Simultaneous estimation of finasteride and tamsulosin hydrochloride by reverse phase HPLC in bulk and pharmaceutical dosage form. International Journal of Pharmacy & Life sciences, 2012 3(8): 1905-1908.
- [20] Chandan, M. Vasudevan World Academy of Science, Engineering and Technology International Journal of Medical, Health,

Pharmaceutical and Biomedical Engineering Vol:7
No:12, 2013

- [21] P. Ravisankar, G. Devala Rao et al Asian journal of pharmaceutical and clinical research Vol 6, Supply 3, 2013 ISSN - 0974-244
- [22] Kullai Reddy Ulavapalli et al Indian Journal of Novel Drug delivery 3(2), Apr-Jun, 2011, 134-142
- [23] Anil Waldiaa, Shubash Guptab et al Journal of Chemical Technology Vol. 15, November 2008, pp. 617-620
- [24] Shreya R. Shah, S. Dey, et al Journal of Taibah University for Science, Volume 8, Issue 1, January 2014, Pages 54-63
- [25] Elena Gabriela Oltean, A. Nica et al Publish in OA Lib Journal ISSN: 2333-9721
- [26] Peyami Sari, Majid Razzak Journal of Liquid Chromatography & Related Technologies Volume 27, Issue 2, 2005